


## Section 1


### Contents

1	Layout of control panel and display .....	1
2	Control panel functions and error messages .....	2

1 Layout of control panel and display



2 Control panel functions and error messages




**Display lights**

- All actual and nominal machine-parameters corresponding. **Film feed is only possible in this case.**

**Display dark**

- error detected
- film in the feed section

- indication of the developer-temperature in the normal case and during heating up
- indication of the fixer-temperature during heating up, if nominal temperature cannot be met. In this case temperature-display is flashing alternating with error 7 indication.



- display for service-programm

**Level monitoring machine tanks**

- intermediate-/main wash by standard
- developer/fixer -> accessory

**LED flashing**

- -> level fault IW/MW/D/F
- actual and nominal developer temperature not corresponding after 30 min.

**Level monitoring replenisher-tanks**

LED flashing

- replenisher-tank empty -> refill tank

**level monitoring disposal-tanks (accessory)**

LED flashing

- Tank is full -> empty tank



A "buzzer" which indicates a brief acoustic signal that further film feed is enabled after the delay time for the previous film feed has elapsed (and the OK display lights up), is available as an accessory.